

WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM

Page 12 of 12

Proficiency Test Report

Lab Id: 11642 USGS NATIONAL WATER QUALITY LABORATORY
 EPA Lab Code: POB 25046-MS 407-BLDG 95
 DENVER, CO-80225-0046
 UNITED STATES
 (303) 236-2000
 Director: MR. MERLE SHOCKEY

Shipment Date : 24-Jan-2005
 Closing Date : 10-Mar-2005
 Score Date : 29-Mar-2005

This report may contain data that are not covered by the NVLAP accreditation.

** Indicates NVLAP accredited analyte evaluated using the USEPA's National Standards for Water Proficiency Testing Studies Criteria Document.
 NVLAP Lab Code 200387-0. ELAP is an A2LA accredited Proficiency Testing Provider. Certificate Number 1785.01

Shipment: 280 Non Potable Water Chemistry

Analyte Name	Units	Sample ID	Method	Result	Mean/ Target *	Warning Limits	Acceptance Limits	Score
Sample: Non Potable Water Purgeables								
trans-1,3-Dichloropropene EPA Code: N/A	ug/L	8013	USGS O-4127-96	164	150		50.6 - 249	Satisfactory <i>116 passed out of 116 reported results.</i>
Tetrachloroethene ** EPA Code: N/A	ug/L	8013	USGS O-4127-96	43.4	41.2 *	31.5 - 48.6	27.2 - 52.9	Satisfactory <i>116 passed out of 121 reported results.</i>
Toluene ** EPA Code: N/A	ug/L	8013	USGS O-4127-96	59.3	52.5 *	42.4 - 60.9	37.8 - 65.5	Satisfactory <i>125 passed out of 129 reported results.</i>
Trichloroethene ** EPA Code: N/A	ug/L	8013	USGS O-4127-96	32.6	29.5 *	22.4 - 35.1	19.2 - 38.2	Satisfactory <i>118 passed out of 123 reported results.</i>
Total Xylenes EPA Code: N/A	ug/L	8013	USGS O-4127-96	404	408		247 - 568	Satisfactory <i>122 passed out of 127 reported results.</i>
Vinyl chloride EPA Code: N/A	ug/L	8013	USGS O-4127-96	< 0.09	< 10.0 *			Satisfactory <i>115 passed out of 116 reported results.</i>
Sample: Non Potable Water Residue								
Solids, Total EPA Code: N/A	mg/L	8002	USGS I-3750-85	355	372		331 - 413	Satisfactory <i>93 passed out of 108 reported results.</i>
Sample: Non Potable Water Total Dissolved Solids								
Solids, Total Dissolved ** EPA Code: N/A	mg/L	8040	USGS I-1750-85	175	178	145 - 211	128 - 228	Satisfactory <i>98 passed out of 107 reported results.</i>